

**PROCUREMENT PROCEDURE OF CPRI (NON WORKS)**

Revision : 04  
 Dt of Rev: 27.08.2020  
 Page No. : 1 of 1  
 Section : Formats  
 Topic : Technical Specifications format

Issue No : 2  
 Issue Dt. : 30.06.2003  
 Issued by : P A  
 Documents : PPM

FORMAT NO.:CPRI/PUR/CTBID/GTP

**Section IV T -Technical Specification**

**CENTRAL POWER RESEARCH INSTITUTE, BENGALURU/BHOPAL Web: www.cpri.res.in**

Tender Enquiry No : CPRI/BLR23SCL35M1187

Description of the Equipment/Goods/Services : Repair of 3Channel WT3000 Yokogawa make Power Analyzer

Note : 1) The technical bid submitted in other than this format is liable to be rejected.

2) All blue fields are mandatorily to be filled in.

Name and address of the bidder					
Quotation Number and Date					
<b>To be completed by the Bidder</b>					
Sl.No.	Technical Specifications/Parameters		Details of guaranteed technical parameters offered by the bidder	Guaranteed Technical Particulars (GTP)	Deviations from GTP
1	Place where equipment /service to be supplied /provided: Short-circuit Laboratory, CPRI, Bengaluru				
2	Scope: Repair & calibration				
3	Quantity	1 no.			
4	Application: The instrument is used for measurement of losses in Distribution Transformers	1 no.			
5	Scope: Scope includes repair and calibration				
6	Nature of repair: The instrument is presently not working. All the three channels are damaged and needs to be repaired by changing the components/cards and updating the firmware if required.				
7	Acceptance criteria: The instrument shall be calibrated at NABL accredited calibration lab and all the errors shall be within the specified permissible limits as claimed by the manufacturer.				
8	Warranty: One year from the date of repair and supply				

PN: 1) Mere statement of "Complied" do not suffice the requirement. The details of technical parameters in proof of CPRI requirements shall be furnished along with technical write-up, catalogues, brochures, literatures, phamplates, or any other documents shall be submitted in hard copy along with technical bid.  
 2) Calibration reports/certificates, factory test reports/certificates from an accredited agencies/facilities shall be submitted wherever applicable.  
 3) CPRI reserves the right to conduct "pre-dispatch inspection" prior to dispatch at the works of the supplier and the expenditure towards PDI shall be borne by CPRI. However information regarding the rediness of the equipment/machinery for the PDI shall be communicated in writing at least 70 days in advance.